

Project-title and REK nr:

2011/133-2 En sammenlignende studie av to prinsipper ved hofteprotesekirurgi (MPC-studien). (A comparison of two principles in THA surgery).

Project-period: 2012-2026

Introduction:

A randomised, comparative, prospective, single-blinded trial of two hip prostheses.

1. Marathon cemented all-poly acetabular cup (DePuy, Warsaw, Indiana), Corail uncemented femoral stem (DePuy, Warsaw, Indiana), 32mm Alumine BioloX forte femoral head (DePuy, Warsaw, Indiana)
2. Pinnacle / Marathon uncemented cup (DePuy, Warsaw, Indiana), Corail uncemented femoral stem (DePuy, Warsaw, Indiana), 32mm Alumine BioloX forte femoral head (DePuy, Warsaw, Indiana)

Material and method:

In each group, 17 patients will receive the Corail stem without a collar and 17 with a collar.

Radiostereometric analysis (RSA) will be used to measure polyethylene wear and component migration. Dual-energy x-ray absorptiometry (DXA) measurements will be used to measure bone mineral density. Furthermore the outcome will be systematically measured by clinical scoring and plain x-rays.

The aims of the study are:

- to compare polyethylene wear and migration of the moderately cross-linked Marathon polyethylene in a cemented version (Marathon) and in an uncemented cup (Pinnacle / Marathon)
- to evaluate micromigration and periacetabular bone mineral density of the two cups.
- to evaluate the migration patterns and proximal femoral bone density changes around Corail stems with and without a collar.

Results:

The two versions of the Corail stem both are sufficiently stable in the femur, but some cases with the collarless version have more initial migration.

Time horizon:

Status 22.12.2023: Two manuscripts are under preparation with 5 year RSA and DXA results. At the end of the study in 2026 another two will be written. After that the study ends.

Contact information:

If you have any questions, please contact the project manager geir.hallan@helse-bergen.no.